

Aiken Technical College
Radiation Protection Technology Program
Air Sample Data Sheet

(For Training Purposes Only—No Actual Hazards Exist)



Location: Room 509 outside A-509-B

Date:
today

RP Technician (s):
JOHN AVERY *J. Avery*

Sampler Information

Model: STAPLEX
TFIA

Serial Number: 23899R

Media: filter
paper

Calibration Due Date: 12-31-11

Sampling Information

1 cubic foot = 2.8316 E⁵ cubic centimeters

Sampling Time		Sample Flow Rate (CFM)		Sample Type: <i>particulate</i>
ON	OFF	ON	OFF	
1015	1025	25	25	Total Volume (cc): 7.08 E6

Counting Data

<i>β</i>	Model: <i>Wdium</i> Model 12 w/ SAV489-110C	Serial Number: CMC009663
	Background (cpm): 100	Calibration Due Date: 1-27-12
<i>α</i>	Model: <i>Wdium</i> Model 12 w/ AC3-7MNV-LV	Serial Number: CMC009674
	Background (cpm): 0	Calibration Due Date: 1-27-12

Type	Counter Efficiency cpm/dpm	Background Count Rate cpm	Quick Count net cpm	6 Hour Count Rate net cpm	24 Hour Count Rate net cpm
Alpha	0.4	0	0		
Beta	0.1	100	4500		

$$\text{Activity } \frac{\mu\text{Ci}}{\text{cc}} = \frac{\text{cpm}}{\text{eff}(\frac{\text{cpm}}{\text{dpm}}) * \text{volume}(\text{cc}) * 0.99 * 2.22\text{E}^6(\frac{\text{dpm}}{\mu\text{Ci}})}$$

$$\alpha \frac{\mu\text{Ci}}{\text{cc}} = \underline{\text{ND}}$$

$$\beta \frac{\mu\text{Ci}}{\text{cc}} = \underline{2.9\text{E}^{-9} \mu\text{Ci}/\text{cc}}$$